

SCIENCE APPLICATIONS INTERNATIONAL CORPORATION
Organic Data Review Checklist - Standard Validation

Project: Harley-Davidson

Page 1 of 11

SDG No: 180-43134-1

Analysis: See attached

Method: See attached

Laboratory: TestAmerica Pittsburgh

Matrix: Water

The above data package has been reviewed and the analytical quality control/quality assurance performance data have been summarized. The general criteria used to assess the analytical integrity of the data were based on an examination of the following:

- Case Narrative
- Analytical Holding Times
- Sample Preservation

Method and Project Blanks

Project Specific QA/QC or contract requirements may take priority over validation criteria in this procedure.

Overall Remarks: No major issues

Definition of Qualifiers:

- "U", not detected at the associated level
- "UJ", not detected and associated value estimated
- "J", associated value estimated
- "R", associated value unusable or analyte identity unfounded
- "=", compound properly identified and value positive

Reviewed by:  Alan G. Miller Jr.

Date: 6/23/15

QA Reviewed by: 

Date: 6-23-15.

AGM
6/23/15

I. Case Narrative

Verify direct statements made within the Laboratory Case Narrative (note discrepancies).

Remarks: *No issues*

II. Re-analysis and Secondary Dilutions

Verify that re-analysis and secondary dilutions were performed and reported as necessary. Determine appropriate results to report.

Remarks: _____

III. Holding Times

Metals - Waters - preserved to pH<2, 180 days from sample collection

Metals - Soils - 180 days from sample collection

Mercury - Waters - preserved to pH<2, 28 days from sample collection

Mercury - Soils - 28 days from sample collection

Deviations:

Sample #	Metals				Mercury			
	Date Collected	Date Analyzed	Days >HT	pH Check	Date Collected	Date Analyzed	Days >HT	pH Check

Actions:

1. If preserved samples exceed holding time, qualify all associated results as estimated (J/UJ).
2. If unpreserved samples exceed holding time, qualify all associated results as unusable (R).
3. If holding times are exceeded by more than 2X, reviewer may qualify non-detected results as unusable (R)
4. If water samples are not acidified, use professional judgement. Minimally, qualify data as estimated (J) and non-detects unusable (R).
5. If soil samples exceed holding time, use professional judgement to qualify data.

Remarks:

NA

III. Holding Times

Sample should be preserved and analyzed according to the appropriate analytical method

In general the following preservations and holding times for waters can be applied:

- Sulfate, 4 degrees C, 28 days
- Sulfide, 4 degrees C, pH \geq 9 with zinc acetate/sodium hydroxide, 7 days
- Bromide/Chloride/Fluoride, no preservative required, 28 days
- Nitrate/Nitrite or Ammonia, 4 degrees C, pH \leq 2 with sulfuric acid, 28 days
- Nitrate or Nitrite, 4 degrees C, 48 hours
- Alkalinity, 4 degrees C, 14 days
- TDS/TSS, 4degrees C, 7 days
- Phosphate (total), 4 degrees C, pH < 2 with sulfuric acid, 28 days
- Hexavalent Chromium, Cool 4 degress C, water- 24 hours, soil - 30 days

Deviations:

Sample #	Analyte	Date Collected	Date Extracted	Date Analyzed	Notes:

Actions:

1. If holding times are exceeded, all results are qualified as estimated (J/UJ)
2. If holding times are exceeded by more than 2X, reviewer may qualify non-detected results as unusable (R)
3. If samples were not properly preserved, use professional judgement to qualify the data

Remarks: WA

VI. Blanks

All blanks were reported per matrix per concentration level for each 12 hour period on each GC/ MS system used to analyze VOCs and SVOCs Yes No

Review associated laboratory and project blank samples. List documented contamination below:

Laboratory Method Blanks:

<u>Date:</u>	<u>Lab ID #</u>	<u>Fraction</u>	<u>Compound</u>	<u>Conc. (ppb)</u>
4/20/15	MB 180-729024/12	VOC	Methylene chloride	0.363

Associated Project Blanks (e.g., equipment rinsates, trip blanks, etc.)

<u>Date</u>	<u>Lab ID #</u>	<u>Fraction</u>	<u>Compound</u>	<u>Conc. (ppb)</u>

Remarks: See above

VI. Blanks (continued)

Calculate action levels based on 10X the highest blank concentration of "common laboratory solvents", VOCs (methylene chloride, acetone, toluene, 2-butanone, cyclohexane) or SVOCs (phthalates), and 5X the highest blank concentration for all other VOC, SVOC, Pesticides, and PCB compounds. Sample weights, volumes, and dilution factors must be taken into account when applying the 5X and 10X criteria. This allows the total amount of contaminant present to be considered.

Deviations:

Compound	Maximum Conc. Detected, (ppb)	Action Level (ppb)	Samples Affected
<i>mB Methylene chloride</i>	<i>0.363</i>	<i>3.63</i>	<i># + None</i>

Actions:

1. If compound results exceed the action levels, the data are not qualified
2. If compound results are below the required reporting level, report results as non-detect (U) at the reporting level
3. If the compound is detected above the reporting level, but below the action level, qualify as not-detected (U)
4. If gross contamination exists in blanks (i.e., saturated peaks by GC/MS), all affected compounds in the associated samples should be qualified as unusable (R) due to interference.
5. If blanks were not analyzed per matrix per concentration level for each 12 hour period on each GC/MS system used to analyze VOCs and SVOCs use professional judgement to qualify data. Data may be rejected (R).

Remarks:

The lot of MC is in the dMB and only applies to the analysis run on 4/20/15. Since this is the sample run that is impacted by the grosser diluted run the result is ignored. So no input.

Hold Time Summary

Sample Number	Sample Name	Method	Date Collected	Analysis Date	Date Extracted	Days to Analysis
180-43134-1	HD-MW-64D-0/1-0	SW846 8260C	4/14/2015	4/20/2015		6
180-43134-1	HD-MW-64D-0/1-0	SW846 8260C	4/14/2015	4/21/2015		7
180-43134-2	HD-MW-141A-0/1-0	SW846 8260C	4/15/2015	4/20/2015		5
180-43134-3	HD-QC1-0/1-2	SW846 8260C	4/15/2015	4/20/2015		5

Trip Blank Detections

Sample ID	Sample	Analyte	Result	5x	10x	Method	Units	Qual
MB 180-139024/12	MB 180-139024/12	Methylene Chloride	0.363	1.815	3.63	SW846 8260C	ug/L	J

Sample	Lab_Id	S_Date	L_Date	Time_Ana	Matrix	Method	Res_Type	CAS	Analyte	Result	Units	Qual	Det_Lmt	Rep_Lmt
HD-MW-64D-0/1-0	180-43134-1	4/14/2015	4/20/2015	19:18	WATER	SW846 8260C	REG	75-09-2	Methylene Chloride	25	ug/L	B	3.1	25
HD-MW-64D-0/1-0	180-43134-1	4/14/2015	4/21/2015	22:16	WATER	SW846 8260C	REG	75-09-2	Methylene Chloride	1.8	ug/L	J	0.31	2.5
HD-MW-141A-0/1-0	180-43134-2	4/15/2015	4/20/2015	20:06	WATER	SW846 8260C	REG	75-09-2	Methylene Chloride		ug/L	U	0.13	1
HD-OC1-0/1-2	180-43134-3	4/15/2015	4/20/2015	19:42	WATER	SW846 8260C	REG	75-09-2	Methylene Chloride		ug/L	U	0.13	1
MB 180-139024/12	MB 180-139024/12		4/20/2015	11:28	WATER	SW846 8260C	BLK	75-09-2	Methylene Chloride	0.363	ug/L	J	0.13	1
MB 180-139148/6	MB 180-139148/6		4/21/2015	12:50	WATER	SW846 8260C	BLK	75-09-2	Methylene Chloride		ug/L	U	0.13	1